



7445 Dean Martin Dr., Suite 107
Las Vegas, Nevada 89139

Certificate of Processing

11/08/2021 10:18:21
 GMT Page 1 of 1

Processing Facility: Ontario

Lot # SS311
 Work Order # 11113-6227
 Sales Order # A A 2334729

SCCI - Gamma Treatment

Irradiation Date/Time: 11/08/2021 10:18 GMT

Quality Test Summary

-----Signed By -----
 User

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
430.00	Cycle Time	1 SEC	1000 SEC	303 SEC	Pass	R. Rodriguez	11/08/2021 10:18:29 GMT
		Reason Code Test					
450.00	Maximum Dose	5.0 kGy	50.0 kGy	35.9 KGY	Pass		11/08/2021 05:57:35 GMT
		Reason Code Test					
450.00	Minimum Dose	50.0 kGy	50.0 kGy	23.0 KGY	Pass		11/08/2021 14:39:59 GMT
		Reason Code Test					

Date: 11/08/2021 10:18:57 GMT

ISO 9001 and ISO 13485 Registered